


VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument


State of Alaska


Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	130299
Name		Christopher M Babcock	ID# 3867 Date 2/3/11 4/15/11
A	Agency	King Cove Dept. of Public Safety	Phone # 497-2210
Instrument Location		Squad Room	
Alco S/N		3A	Target Value .080 High Pressure 1100
B	Alco Test Values	.078 1 st Alco	.077 2 nd Alco
Signature			
(OVER)			


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.


Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 15 day of JUNE, 2010. 


(Notary Seal Stamp)
Daniel L. Nicholas
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130299

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130299

APRIL 15, 2011

CRF 4/22/11

OPERATOR'S NAME:
BABCOCK CHRISTOPHER M
OPERATOR'S NUMBER: 3867
SUBJECT'S LAST NAME:
UOC
SUBJECT'S FIRST NAME/MI :
UOC
O.L. #: 123456
DEPT/AGENCY: KUC1
CASE/REPORT: UOC
TEST TYPE: U
ALCO TARGET VALUE: .000
ALCO S/N: 3A

— BREATH ANALYSIS —

.000 ADJUSTED FOR 30.32 in
ALCO TARGET .001 11:40
BLANK TEST .000 11:41
INTERNAL STANDARD VERIFIED 11:41
ALCO TV 30.32 in .078 11:41
BLANK TEST .000 11:42
SUBJECT SAMPLE .000 11:43
BLANK TEST .000 11:44
ALCO TV 30.32 in .077 11:44
BLANK TEST .000 11:45

CRF 4/22/11

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130299

APRIL 15, 2011
TIME 11:27

— DIAGNOSTIC CHECK —

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 42c
BAROMETER: 30.32 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

"0000"()**.-./0123456789:;<=>?@ABCDEF6
HIJKLMNPOQRSTUVWXYZ\]^_`'abcde{ghijklmno
pqrstuvwxyz{|}~■